aeroqual

Check particle profiler for leaks

Written By: Tanya Taylor



INTRODUCTION

A leak can be caused by loose connections, worn seals (o-rings) or split tubing.

If a leak occurs, this can affect the flow rate which will have a big impact on the accuracy of the measurement.

To understand how often you should perform this service activity, click here.

PARTS:

- Leak gauge (1)
- Luer fittings (1)

Step 1 — Enter service mode

RH %
1.00
0.0
RH %
86.
86.0
86.6
87.6
87.6

 Enter service mode so any fluctuations in the data caused from this activity can be excluded from air quality reports.

Step 2 — Open sample & block purge



- Adjust the sample flow by pulling the adjustment knob outwards, turning the knob to increase/decrease flow, and pushing the knob back in to lock when desired flow has been reached.
- Fully open the sample flow adjustment valve, which is located on the face plate of the pump module.
 - On older monitors, the sample flow adjustment value is positioned at the bottom of the enclosure under the PDI cover.
- Block the purge flow by pinching the purge tube.
 - (i) On older monitors, block the purge flow by disconnecting the purge line and capping the purge line and ports with luer caps.

Step 3 — Attach gauge



- Remove the TSP inlet and place the vacuum gauge on the inlet.
- Wait for the gauge to stabilise.
- (i) The gauge should reach at least -60 kPa while power is connected.

Step 4 — Observe pressure change



- Stop the sample pump by pulling out the black and red power cable from the electronics module.
- Count how long it takes for the pressure to change by 10 kPa.
- (i) If the pressure change (leak rate) is greater than 10 kPa in 10 seconds, you'll need to check for loose connections, worn seals (o-rings) or split tubing.
- *(i)* For example, if the needle moves from -70 kPa to -60 kPa in 20 seconds, this is OK, but if if moves from -70 kPa to -60 Kpa in 8 seconds, this indicates a leak.

Step 5 — Record in journal

All journal types V User entry Cloud user - John Wagner			
	Instrument in good condition	Cooling fan operational	
	No obstructions to monitoring equipment	PM and gas inlet secure	
8. Equipment:		Instrument has been running at stab	
Aeroqual Gas dilutio	on calibrator: Aircal 1000		
Aeroqual Ozone cal	ibrator: AQM O3Cal		
Aeroqual Flow meter	er AQM R7	4 Gas cylinders:	
		CO 1000 ppm in Air (expiry Mar	
		SO2 20 ppm in Air (expiry Dec	
		NO2 20 ppm in Air (expiry Nov	
4. Flow rate check:	Expected flow rate = 0.450 ml per min,		
Measured flow rate = 0.452 ml per min		Open door and change gas inlet filt	
Main inlet flow rate	OK, individual module flow rates were not measured.		
6. Zero calibration			
All modules passed	zero calibration, all modules were stable and all offsets we	vere within acceptable limits.	
7. Span Calibration			
CO @ 10.00 pm	Module response was 8.95 ppm gain adjustment to 1.1.	15 pass	
SO2 @ 0.2 ppm	Module response was 0.210 ppm gain adjustment to 0.1	0.92 pass	
NO2 @ 0.2 npm	Module response was 0.090 ppm gain adjustment to 2.1	10 pass (module may need replacing soon contact.	

- Record the results of this service activity in the monitor's journal.
- Exit service mode.

For further support, contact <u>Technical Support</u>.