aeroqual

Zero calibration check for PCX

Written By: Kyle Alberti



INTRODUCTION

Use this procedure to check the zero baseline of your PCX.

The PCX has an automatic zero calibration function which runs at start up and every 24 hour after that, this procedure automatically changes the zero offset of the PCX.

To understand how often you should perform this service activity, click here.

PARTS:

• Zero filter & flow assembly PCX (1)

Step 1 — Enter service mode

Normal operation									
Normal operation									
Calibration	Manual se	Manual service mode Start							
History									
Manual Entry	Calibratic	on parar	neters						
Manual Lifery		NO2	Ox	O3	O3 raw	PM2.5 raw	PM2.5	TEMP	RH %
	Gain	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.00
	Offset	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.0
	а	1.100		2.550					
	b			1.870					
	Real time								
	Time	NO2 ppb	Ox ppb	O3 ppb	O3 raw	PM2.5 raw µg/m³	PM2.5 µg/m ³	TEMP °C	RH %
	1 <mark>1:4</mark> 2 a.m.	2.9	29.6	24.2	23.7	1.7	1.1	15.74	86.8
	11:41 a.m.	2.8	29.2	24.0	23.5	1.6	1.0	15.63	86.0
	11:40 a.m.	3.1	29.7	24.2	23.8	1.9	1.2	15.60	86.6
	11:39 a.m.	3.6	30.2	24.1	23.7	1.5	1.0	15.55	87.
	11:38 a.m.	4.7	30.4	23.4	23.0	1.3	0.8	15.48	87.6
		and the second second							1.1.1.1.1.1.1

 Enter service mode so any fluctuations in the data caused from this activity can be excluded from air quality reports.

Step 2 — Attach zero filter



- Remove the TSP head and O-rings from the top of your monitor's particle inlet.
- Attach the flow adaptor and zero filter.

Step 3 — Check concentrations

00							001				
Normal operation							Calibration and S	ervice 🗸 Instrument	~		
Diagnostics	Select para	meter PM10	Averaging pe	riod 1 minute	✓ Pause	Export	Normal operation				
Download Data				_			Calibration				
Madula Davida	Time	PM10 (µg/m³)	Run time (Hours)	Pressure (mBar)	Laser current (mA)	Raw signal (Hz)		Manual service mod	ie St	art	
Module Details	11:42 PM	1.39	7663.000	1017.000	20.800	101.583	History				
Module Settings	11:41 PM	1.39	7663.000	1017.000	20.800	101.833		Calibration paran	neters		
, in the second s	11:40 PM	1.39	7663.000	1017.000	20.800	104.000	Manual Entry 🔰		01410		
	11:39 PM	1.38	7663.000	1017.000	20.800	102.667	7 011 1		µg/m ³		
	11:38 PM	1.38	7663.000	1017.000	20.800	101.667	Zero Calibration	Gain	1.000		
	11:37 PM	1.29	7663.000	1017.000	20.800	103.667		Goin			
	11:36 PM	1.48	7663.000	1017.000	20.800	99.667		Offset	0.00		
	11:35 PM	1.36	7663.000	1017.000	20.800	107.167		Deal time measure			
	11:34 PM	1.33	7663.000	1017.000	20.800	103.333		Real time measur	ements	Last 5 readings 💙	
	11:33 PM	1.38	7663.000	1017.000	20.800	101.833			PM10		
	11:32 PM	1.40	7663.000	1017.000	20.800	103.167		Time	µg/m°	Inlet	
	11:31 PM	1.42	7663.000	1017.000	20.800	100.167		11:41 PM	1.39	Sample	
	11:30 PM	1.46	7663.000	1017.000	20.800	103.167		11:40 PM	1.39	Sample	
	11:29 PM	1.41	7663.000	1017.000	20.800	101.500		11-20 PM	1 2 9	Sample	
	11:28 PM	1.50	7663.000	1017.000	20.800	102.000		11.33 FIVE	1.50	Sample	
	11:27 PM	1.48	7663.000	1017.000	20.800	102.417		11:38 PM	1.38	Sample	
	11:26 PM	1.42	7663.000	1017.000	20.800	101.500		11:37 PM	1.29	Sample	
	11:25 PM	1.53	7663.000	1017.000	20.800	101.500		Average	1.37		
	11:24 PM	1.49	7663.000	1017.000	20.800	103.167		Std Day	0.04		
	11:23 PM	1.34	7663.000	1017.000	20.800	102.750		Std Dev	0.04		
	11:22 PM	1.54	7663.000	1017.000	20.800	103.083					
	11:21 PM	1.46	7663.000	1017.080	20.800	102.583					
	11-20 DM	1 /7	7663.000	1017 170	20.800	102 500					

- To view particle concentrations, open the Diagnostics and Advanced app and select Diagnostics from the side menu.
- Alternatively, use the **Calibration and Service** app and select **Manual Entry** from the side menu.
- While the zero filter is attached, the PM μg/m3 values for PCX channels should be 0 or close to 0 (between ±3 μg/m3).

Step 4 — Adjust offset

Calibration and	Service ~ Instrument ~ DOV	VNWIND (DS-VIRTUAL-02) 🗸 Journal 🗸 Public Dem
Normal operation		
Calibration	> System Particles >	ΡΜ10 (μg/m³)
Manual Entry Zero Calibration	Current Average Std PM10 3.08 2.64 0.21 Average and standard deviation are calculated from last 10 minutes 100.00 100.00 50.00 0.00 100.00 100.00	50.00
		23:50 00:00 00:10 00:20 00 Start calibration run Offset Gain
		Current offset New o 0.00 Recommend

- If the zero check shows negative numbers, there may be a problem with the auto zero cycle. See the troubleshooting section.
- Check that the offset is 0.000
- Check that the gain is between 0.6 and 4.0
- The zero baseline check is now complete.

Step 5 — Record in journal

All journal types	1							
Jser entry Cloud use	r · John Wagner							
1. Site Inspection:	No new local emission sources		2. Instrument inspection:					
	Instrument in good condition		Cooling fan operational					
	No obstructions to monitoring equipment		PM and gas inlet secure					
3. Equipment:			Instrument has been running at stab					
Aeroqual Gas diluti	on calibrator: Aircal 1000							
Aeroqual Ozone ca	librator: AQM O3Cal							
Aeroqual Flow meter	er AQM R7		4 Gas cylinders:					
			CO 1000 ppm in Air (expiry Mar					
			SO2 20 ppm in Air (expiry Dece					
			NO2 20 ppm in Air (expiry Nove					
4. Flow rate check:	Expected flow rate = 0.450 ml per min,							
	Measured flow rate = 0.452 ml per min		Open door and change gas inlet filt					
Main inlet flow rate	OK, individual module flow rates were not measured.							
6. Zero calibration								
All modules passed	zero calibration, all modules were stable and all offsets	were within	n acceptable limits.					
7. Span Calibration								
CO @ 10.00 pm	Module response was 8.95 ppm gain adjustment to	1.15 pass						
SO2 @ 0.2 ppm	Module response was 0.210 ppm gain adjustment to	0.92 pass						
NO2 @ 0.2 ppm	Module response was 0.090 ppm gain adjustment to	2.10 pass	(module may need replacing soon contact /					

- Record the results of this service activity in the monitor's journal.
- Exit service mode.

For further support, contact <u>Technical Support</u>.