



Clean PCX TSP inlet

Written By: Kyle Alberti



INTRODUCTION

The PCX uses a TSP inlet to protect the monitor from water ingress. This part can be taken apart and cleaned for optimal performance.

To understand how often you should perform this service activity, [click here](#).



TOOLS:

- [Can of compressed air](#) (1)
 - [Lint-free wipes](#) (1)
-

Step 1 — Enter service mode

oo

Calibration and Service

Instrument

Sales & Support Demo AQY (AQY Demo-001)

Normal operation

Calibration

History

Manual Entry

Manual service mode

Start

Calibration parameters

	NO2 ppb	Ox ppb	O3 ppb	O3 raw ppb	PM2.5 raw µg/m³	PM2.5 µg/m³	TEMP °C	RH %
Gain	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000
Offset	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.0
a	1.100		2.550					
b			1.870					

Real time measurements

Time	NO2 ppb	Ox ppb	O3 ppb	O3 raw ppb	PM2.5 raw µg/m³	PM2.5 µg/m³	TEMP °C	RH %
11:42 a.m.	2.9	29.6	24.2	23.7	1.7	1.1	15.74	86.1
11:41 a.m.	2.8	29.2	24.0	23.5	1.6	1.0	15.63	86.1
11:40 a.m.	3.1	29.7	24.2	23.8	1.9	1.2	15.60	86.1
11:39 a.m.	3.6	30.2	24.1	23.7	1.5	1.0	15.55	87.1
11:38 a.m.	4.7	30.4	23.4	23.0	1.3	0.8	15.48	87.1

- [Enter service mode](#) so any fluctuations in the data caused from this activity can be excluded from air quality reports.

Step 2 — Dissassemble



- Loosen the three screws on top of the inlet to disassemble the TSP inlet
- *i* The three screws do not need to be removed completely in order to take the base off and will be retained by self-retaining spacers.
- Take the base off
- Remove the insect mesh

Step 3 — Clean TSP inlet



- Clean the pieces with compressed air and a lint-free cloth

Step 4 — Reassemble



- Attach the insect mesh.
- Take the base off
- Tighten the three screws on top of the inlet. The screws should each be tightened gradually in a circle in order to ensure they all thread in properly.

Step 5 — Record in journal

Instrument

Air Quality Monitor (AQM65 04082015-437)

All journal types

User entry | Cloud user · John Wagner

1. Site Inspection:	No new local emission sources Instrument in good condition No obstructions to monitoring equipment	2. Instrument inspection:	Cooling fan operational PM and gas inlet secure Instrument has been running at stable
3. Equipment:	Aeroqual Gas dilution calibrator: Aircal 1000 Aeroqual Ozone calibrator: AQM O3Cal Aeroqual Flow meter: AQM R7	4 Gas cylinders:	CO 1000 ppm in Air (expiry March) SO2 20 ppm in Air (expiry December) NO2 20 ppm in Air (expiry November)
4. Flow rate check:	Expected flow rate = 0.450 ml per min, Measured flow rate = 0.452 ml per min Main inlet flow rate OK, individual module flow rates were not measured.	5. Open door and change gas inlet filter	
6. Zero calibration	All modules passed zero calibration, all modules were stable and all offsets were within acceptable limits.		
7. Span Calibration	CO @ 10.00 ppm Module response was 8.95 ppm gain adjustment to 1.15 pass SO2 @ 0.2 ppm Module response was 0.210 ppm gain adjustment to 0.92 pass NO2 @ 0.2 ppm Module response was 0.090 ppm gain adjustment to 2.10 pass (module may need replacing soon contact Aeroqual)		
8 Pack up. Next scheduled calibration 3 months from now. June 2017.			

- [Record the results of this service activity in the monitor's journal.](#)
- [Exit service mode.](#)

For further support, contact [Technical Support](#).