## aeroqual

# **Configure monitor**

Written By: StJohn Vuetilovoni



## INTRODUCTION

Use this guide to change the monitor ID, location ID, concentration unit, logging frequency, output sensor, alarm settings, control settings, and date and time.

**Note**: It assumes you have already <u>installed the monitor's configuration software</u>, <u>downloaded the</u> <u>cable driver</u> and <u>connected your monitor to a PC</u>.

#### Step 1 — Update monitor settings



- Launch the Aeroqual S500 software and click Monitor > Setup from the menu bar. The Monitor Setup dialog box appears.
- Type in, or use the up/down buttons to set desired values for monitor ID, location ID, concentration unit, logging frequency, output sensor, alarm settings, and control settings.
- If the temperature and relative humidity sensor is connected, an option to select the temperature unit also appears.
- (i) Each function has a keyboard shortcut. The shortcut keys are noted alongside each function in the drop-down menus.

### Step 2 — Set date and time

Ø Monitor Setup	×	
Configuration		
Monitor ID	1	
Location ID	1 🖨	
Concentration Unit:	ppm 🗸	
Logging frequency (Minutes):	5 🔹	
Output sensor	CO2 ~	
Alarm Settings		
Alarm High:	0.3 🜩	
Alarm Low:	0.1 🖨	
Set Alarm Permanent Mute		
Control Settings		
Control High :	0.3 🖨	
Control Low :	0.1	
Set key mute		
Read Save	Update Clock	
ОКС	ancel	

For further support, contact <u>Technical Support</u>.

- To log data, it's essential the clock on the monitor is set to the true time. Click Update Clock to set the date and time. Note: This can also be done on the monitor via its setup menu.
- Click Save to synchronize the monitor values with those from the PC.
- If the clock doesn't synchronize, the monitor clock battery may be flat and need replacement. In this situation, contact <u>Technical Support</u>.